

MRI Safety Week

Name: _____

Dept/Ext: _____

1. J Athletic apparel with silver or copper microfibers that can cause burns.
2. I Responsible for ensuring MR safe practice guidelines and makes final determination of whether or not to scan a patient by risk assessment.
3. D An object or device that may be safely used in the MR environment, provided the conditions for safe use are met.
4. L An object or device known to present safety risks in the MR environment.
5. C An object or device that is safe in all MR environments, without conditions.
6. E Some drug delivery patches contain metallic components and have the potential to cause _____.
7. N If a fire occurs in MRI, this is used.
8. B Divides the facility into 4 safety zones with access restricted to personnel that have been through MR safety training.
9. G Researching and following FDA recommendations on MRI Safety and compatibility of medical devices.
10. K Remove _____ to prevent heating that could possibly cause burns.

- A. Hearing Protection
- B. MRI Zone Map
- C. MR Safe
- D. MR Conditional
- E. Thermal Injury
- F. Level 2 MR Personnel
- G. Device Screening
- H. Magnet
- I. MRMD
- J. Antimicrobial Clothing
- K. EKG wires/leads
- L. MR Unsafe
- M. Ferromagnetic
- N. MR Conditional Fire Extinguisher



